



2023 COMBOSTIK[®]

product brochure

Urine Reagent Strips
Urine Chemistry analyzer

 **DFI Co.,Ltd.**

388-25, Gomo-ro, Jillye-myeon, Gimhae-si,
Gyeongsangnam-do, 50875, Korea (South)

Tel: +82-55-346-1882, Fax: +82-55-346-1883

E-mail: info@dficare.com

Home Page: www.dficare.com

COMBOSTIK® URINE CHEMISTRY ANALYZER

ComboStik® Urine Chemistry Analyzers are reflectance photometer.

The strip is illuminated by white light and the reflected light from the strip is detected by the sensor.

The RGB signals are digitized and this digitized image is interpreted by the processor.

The intelligent imaging analysis software reads the strip and the pads, and based on these color data the parameter values are determined. The results including the date and time of the measurement, sequence number and ID are stored.

ComboStik® Urine Chemistry Analyzers are easy to use and provide reliable results every time.

ORDERING INFORMATION FOR COMBOSTIK® URINE CHEMISTRY ANALYZER

Product Name	Cat. No	SPECIFICATION
ComboStik® R-50S	5640	45 Tests(Max 120 Tests)/Hour, No Built-in Printer, Available up to 11 parameters
DP-mini	E100	External printer for EXCLUSIVE use with R-50S
ComboStik® R-300	5240	300 Tests/Hour, Built-In Printer, Available up to 13 parameters
ComboStik® R-600S	5340	600 Tests/Hour, LCD Touch Screen, Built-In Printer, PC software, LIS connection etc.

R-600S

SEMI-AUTOMATED MULTIPLE ANALYZER



Technical Specification

Throughput : 600 tests / hour

Memory : 5,000 tests

Optional accessories : Barcode Scanner (USB type)

Display : Colorful 7" touch LCD

Interface : USB, RS-232C

SELF-SYSTEM CHECK

Auto-Calibration

Automated F/W update by USB

Connectivity to LIS (Laboratory Information System)



R-50S

SEMI-AUTOMATED SINGLE-TEST



Technical Specification

Throughput : 45 tests/hour (Max 120 tests/hour)
Power : 12V DC/3.33A or Battery 1.5V x 8EA
Memory : 2,000 tests
Net weight : 460 g
Available up to 11 parameters
Optional item : External printer / Bluetooth

Product Features

Light and Portable analyzer
User friendly designed with compact size
Colorful 4.3 inch touch LCD Display
Possibility with 8-battery 1.5 V
Rapid test result in 1 minute
Auto-Calibration function
Accessible to laboratory information system(LIS)
by PC protocol

R-300

SEMI-AUTOMATED SINGLE-TEST



Technical Specification

Throughput : 300 tests/hour (Max 800 tests/hour)
Memory : 20,000 tests
Net weight : 1.3 kg
Available up to 13 parameters
Optional item : Keyboard / Barcode Reader

Product Features

Perfect size for using in laboratory or clinic
Efficient working cycle with multi-strip tray
Simple update available by USB port
Optional operation mode
(General / One by one / Quick mode)
Easy to clean and use without special training
Accessible to laboratory information system(LIS)
by PC protocol

COMBOSTIK® URINE REAGENT STRIPS




ComboStik® Urine Reagent Strips are dip-and-read test strips for the rapid determination of Blood, Bilirubin, Urobilinogen, Ketones, Glucose, Protein, Micro-albumin, Creatinine, Nitrites, Leukocytes, pH, Specific Gravity and Ascorbic Acid (Vitamin C) in Urine.

User friendly designed ComboStik® Urine Reagent Strips are easy to use for preliminary diagnosis and initial screening of health conditions by doctors, laboratories and also end-users. Test result may provide information regarding the status of carbohydrate metabolism, kidney and liver function, acid-base balance, and urinary tract infection.

ComboStik® Urine Reagent Strips are easy to get test results visually/manually based on the stable color development. It is designed to be resistant to vitamin C interference for excellent and reliable test results.

Specialized ComboStik® Micro-albumin/Creatinine test is used to diagnose Micro-albuminuria (an early marker of Diabetic Nephropathy) by Micro-albumin/Creatinine ratio and a simple test for effective management of kidney disease in patient with diabetes and potentially high risk for kidney problem in early stage.



ORDERING INFORMATION

FOR COMBOSTIK® URINE REAGENT STRIPS

Product Name	Cat. No	Tests	Uro	Glu	Bil	Ket	S.G	Bld	pH	Pro	Nit	Leu	AsA	Ma	Cre
G	2104	100		•											
2GP	2204	100		•						•					
2GK	2214	100		•		•									
3	2304	100		•					•	•					
3GK	2314	100		•		•				•					
4SG	2414	100		•			•		•	•					
10	0054	100	•	•	•	•	•	•	•	•	•	•			
11	0964	100	•	•	•	•	•	•	•	•	•	•	•		
11	0965	150	•	•	•	•	•	•	•	•	•	•	•		
2AC*	2223	50													•
12AC*	2B04	100	•	•	•	•	•	•	•	•	•	•			•
13AC*	2C04	100	•	•	•	•	•	•	•	•	•	•	•		•
2MAC* [Ⓢ]	2233	50													•
12MAC* [Ⓢ]	2B14	100	•	•	•	•	•	•	•	•	•	•			•
10M [Ⓢ]	2004	100	•	•	•	•	•	•	•	•	•	•			
11M [Ⓢ]	2904	100	•	•	•	•	•	•	•	•	•	•	•		

◆ Remark  0197

ComboStik® Urine Reagent Strips with

* Micro-albumin/Creatinine test parameters are included

[Ⓢ] Compensation pad is included

Abbreviation: Uro(Urobilinogen), Glu(Glucose), Bil(Bilirubin), Ket(Ketone), S.G(Specific Gravity), Bld(Blood), pH(pH), Pro(Protein), Nit(Nitrite), Leu(Leukocytes), AsA(Ascorbic Acid), Ma(Microalbumin), Cre(Creatinine)